

Corning Incorporated – Life Sciences

One Becton Circle

Durham, NC 27712 USA

www.corning.com/lifesciences

Refer to website for regional contact information.

PAGE: 1 of 1

Date: March 12, 2014

Corning Incorporated - Life Sciences certifies that Falcon® products meet the following criteria:

Falcon Catalog #:	Lot #:	Irradiation Date:	Irradiation Dosage	
			Min:	Max:
351058	3238165	2013/09/07	12.11 kGy	17.92 kGy

Sterilization - Product labeled as sterile is irradiated and dosimetrically released upon US Association for the Advancement of Medical Instrumentation (AAMI) recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. Falcon products labeled sterile meet a minimum requirement of 10^{-3} SAL (Sterility Assurance Level).

This product was sterilized at the following facility:

Corning Incorporated - Life Sciences

One Becton Circle

Durham, NC 27712 USA

VERIFICATION:



Tara Rogers
Quality Manager