

VWR® TEST TUBE RACK

As per the manufacturer, the below product meets the following criteria:

VWR Catalog Number	89215-762
Lot Number	V211014
Description	VWR RACK 13MM-72-PL NYLON WHT
Date of Manufacture:	October 21, 2014
Country of Origin:	India.

Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Autoclaving	TPPL/TEST/001	General Insp. Level- I, AQL- 1.5	PASS
Dimensional	QSF-QC-02	General Insp. Level- I, AQL- 1.5	Within Tolerance

Product Specifications

Material: 30 % Glass filled Nylon with color masterbatch.

BSE/TSE Statement: The above mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.

Non-Pyrogenic: Not Applicable.

DNase & RNase Free: Not Applicable.

Cytotoxicity: Not Applicable.

Latex Free Statement: No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

Signed:



Ken Crossley
Manager
Quality Assurance

Date: October 21, 2014