



Quality Certificate

VWR® CULTURE TUBES

As per the manufacturer, the below mentioned product meets the following criteria:

North American Catalog No:	53283-685
Lot Number:	24121357
Description:	VWR TUBES CULTURE 20X125 CS500
Date of Manufacture:	July 21, 2016 – July 22, 2016

The product was manufactured using glass tubing that conforms to the ASTM Specifications for Glass in Laboratory Apparatus for:

Type I, Class "B" Borosilicate 51 Expansion Glass – Clear,

Which also conforms to the most current version of the United States Pharmacopoeia (USP 35) for Type I Borosilicate glass and will meet the USP limits for glass grain (powdered durability). It also meets the requirements of the most current revisions of the European pharmacopoeia (EP) and Japanese pharmacopoeia for Type I Hydrolytic Glass.

The above product was inspected and accepted in accordance with the Manufacturer's Quality Management System and corresponding inspection and monitoring procedures to an Acceptance Quality Limit of 1.0 per ANSI/ASQC Z1.4.

Signed:

A handwritten signature in blue ink, appearing to read "Ken Crossley", with a stylized flourish at the end.

Ken Crossley
Manager
Quality Assurance

Date : Nov 21, 2016