

Certificate of Quality

REF

454030

LOT

B180539J

Product description: VACUETTE® TUBE 4 ml NH Sodium Heparin

13x75 green cap-green ring, non-ridged

Following inspections were carried out and passed:

visual passed physical passed chemical passed

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

98/79/EC "Directive 98/79/EC of the European Parliament and of the Council of 27th Oct 1998 on in vitro diagnostic medical devices"

ISO 6710 "Single-use containers for venous blood specimen collection"

ISO 11137 "Sterilization of health care products - Radiation"

EN 14820 "Single-use containers for human venous blood specimen collection"

ISO 9001 "Quality management systems - Requirements"

ISO 13485 "Medical devices - Quality management systems - Requirements for regulatory purposes"

><

2019-09-12

STERILE

SAL 10⁻⁶

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2018-06-26

www.gbo.com