



## Quality Certificate

### VWR\* EXAM GLOVES

This document certifies that the below product meets the following criteria:

North American Catalog No:	82026-426
Lot Number:	234R11B001
Description:	VWR GLOVE NITRILE PF M PK100

**Material :** Nitril BUTADIENE Rubber, 100% Latex Free.

**QC Testing :** Representative production samples are collected and inspected in accordance with current applicable product specifications:  
Meet or exceed ASTM D 6319-00a(Water test & Dimension)  
Meet or exceed EN 455-1&2  
ISO 2859 (Sampling table)

**Accuracy:** Water tight test :Inspection level=G1; AQL 1.5  
Visual Inspection:Inspection level=G1; AQL 1.5(critical defects)  
AQL 2.5(major defects)  
AQL 4.0(minor defects)

**Dimensional:** Inspection level=S2; AQI 4.0  
**Physical property:** Inspection level=S2; AQI 4.0  
**Powder Content:** Inspection level=5 pieces;Average  $\leq$  2mg/glove  
**Moisture Content:** Inspection level=5 pieces;Average  $\leq$  1%  
**Glove Content:** Inspection level=5 dispensers:Tolerance 100+/-2 pieces  
**Sterility:** N/A  
**Non-Pyrogenic:** N/A  
**Rnase/Dnase Free:** N/A  
**Regulatory:** FDA 510K,ISO 13485,CMDCAS Quality Systems,GMP Certificate

Signed:

Ken Crossley  
Manager  
Quality Assurance

Date of Issue: Feb 12, 2011