



## Quality Certificate

### VWR\* GLOVES

This document certifies that the below product meets the following criteria:

<b>North American Catalog No:</b>	82026-424
<b>Lot Number:</b>	PT234R13C004
<b>Description:</b>	VWR GLOVE NITRILE PF S PK100
<b>Manufacturing Date:</b>	March 2013
<b>Product shelf Life:</b>	February 2018

**Material:** Nitrile Butadiene Rubber, 100% latex Free

**QC Testing:** Representative production samples are collected and inspected in accordance with current applicable product specifications.

Meet or exceed ASTM D 6319-00a (Water Test & Dimension)

Meet or exceed EN 455-1&2

ISO 2859 (sampling table)

<b>Accuracy:</b>	<b>Water tight test:</b>	Inspection level = G1; AQL1.5
	<b>Visual inspection:</b>	Inspection level = G1; AQL1.5 (critical defects) AQL2.5 (major defects) AQL4.0 (minor defects)
	<b>Dimensional:</b>	Inspection level = S2; AQL4.0
	<b>Physical Property:</b>	Inspection level = S2; AQL4.0
	<b>Powder Content:</b>	Inspection level = 5pieces; Average $\leq 2\text{mg/glove}$
	<b>Moisture Content:</b>	Inspection level = 5pieces; Average $\leq 1\%$
	<b>Glove Count:</b>	Inspection level = 5 dispensers; Tolerance 100+/-2pieces

**Sterility:** N/A

**Non-Pyrogenic:** N/A

**Rnase/Dnase Free:** N/A

**Regulatory:** FDA510K, ISO 9001, ISO 13485, CMDCAS Quality System, GMP Certificate

Signed:

Ken Crossley  
Manager  
Quality Assurance

Date of Issue: March 31, 2013