



Certificate of Compliance

Product description: VACUETTE® Blood Collection Tube
3 ml Coagulation Sodium Citrate 3.2%

Greiner product No.: 454334
Lot No.: B140611W
Expiry date: June - 2015

Sterility: Sterile

1	Visual checks	☒	passed
2	Physical tests	☒	passed
3	Chemical tests	☒	passed

We hereby certify that the tubes were tested in accordance with the control procedures stated in the following international standards and meet the required specifications:

ISO 6710 "Single-use containers for venous blood specimen collection"
ISO 11137 "Sterilization of health care products – Requirements for validation and routine control – Radiation sterilization"

The product has been manufactured under controlled conditions in accordance with ISO 9001 and EN ISO 13485.



Sept 23, 2012
Date

[Signature]
Carolyne Guthrie
Quality Manager
on behalf of...