



## VWR® POWDER-FREE NITRILE EXAMINATION GLOVES

As per the manufacturer, the below product meets the following criteria:

North American Catalog No.	82026-424
North American Catalog No:	02020-424
Lot No:	PT17P003
Description:	VWR GLOVE NITRILE PF S PK100
Date of Manufacturing:	November 2017
Product Shelf Life:	October 2022
Country of Origin:	Indonesia

## **Quality System Compliance:**

Products are manufactured under FDA510K, ISO 9001:2008, ISO 13485:20031, CMDCAS Quality System standard and Good Manufacturing Practices. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

QC Testing: Representative production samples are collected and inspected in accordance with current

applicable Product specification:

Meet or exceed ASTM D3578 and ISO 2859 (Sampling table)

**Accuracy:** Water tight test: Inspection level = G1; AQL 1.5

Visual inspection: Inspection level = G1; AQL 1.5 (critical defects) AQL 2.5 (major defects)

AQL 4.0 (minor defects)

Dimensional: Inspection level = S2; AQL 4.0 Physical Property: Inspection level = S2; AQL 4.0

Powder Content: Inspection level = 5 pieces; Average ≤ 2 mg/glove

Moisture Content: Inspection level = 5 pieces; Average ≤ 1%

Glove count: Inspection level = 5 dispensers; Tolerance 100+/-2 pieces

Product Specification: This product is Class I category device as defined by the FDA in 21CFR parts 807 sub

parts C

Material: Nitrile Butadiene Rubber

Non-Pyrogenic: N/A

DNase & RNase Free Statement: N/A

Cytotoxicity Statement: N/A

BSE/TSE Statement: N/A

**Latex Statement:** This product is not made with natural rubber latex.

Signed:

Linda K. Shine

Manager - Supplier Qualification

Global Compliance

VWR International LLC Date: November 03, 2017