

Certificate of Quality

REF

456073P

LOT

B18103NU

Product description: VACUETTE® TUBE 5 ml Z Serum Separator Clot Activator
13x100 red cap-yellow ring, non-ridged

Following inspections were carried out and passed:

visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed
chemical	<input checked="" type="checkbox"/>	passed

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

98/79/EC "Directive 98/79/EC of the European Parliament and of the Council of 27th Oct 1998 on in vitro diagnostic medical devices"

ISO 6710 "Single-use containers for venous blood specimen collection"

ISO 11137 "Sterilization of health care products - Radiation"

EN 14820 "Single-use containers for human venous blood specimen collection"

ISO 9001 "Quality management systems - Requirements"

ISO 13485 "Medical devices - Quality management systems - Requirements for regulatory purposes"



2020-02-08

STERILESAL 10⁻⁶

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2018-12-04

www.gbo.com

Austria: Greiner Bio-One GmbH, office@at.gbo.com | **Belgium:** Greiner Bio-One BVBA/SPRL, info@be.gbo.com | **Brazil:** Greiner Bio-One Brasil, office@br.gbo.com
China: Greiner Bio-One Suns Co. Ltd., office@cn.gbo.com | **Germany:** Greiner Bio-One GmbH, info@de.gbo.com | **France:** Greiner Bio-One SAS, infos@fr.gbo.com
Hungary: Greiner Bio-One Hungary Kft., office@hu.gbo.com | **India:** Greiner Bio-One India Pvt. Ltd, info@gboindia.com | **Japan:** Greiner Bio-One Co. Ltd., info@jp.gbo.com
Netherlands: Greiner Bio-One B.V., info@nl.gbo.com | **Switzerland:** Greiner Bio-One Vacuette Schweiz GmbH, office@ch.gbo.com
Thailand: Greiner Bio-One Thailand Ltd., office@th.gbo.com | **UK:** Greiner Bio-One Ltd., info@uk.gbo.com | **USA:** Greiner Bio-One North America Inc., info@us.gbo.com