



VWR® LABORATORY BOTTLES, POLYPROPYLENE, NARROW MOUTH

As per the manufacturer, the below product meets the following criteria:

VWR Catalog Number:	414004-183
Lot Number:	VAV300119
Description:	VWR BOTTLE PP NM 1000ML PK6
Date of Manufacture:	January 30, 2019
Country of Origin:	India.

Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2015 & ISO13485:2016 Quality Management System. Inspection records are reviewed and signed for product release.

QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
Leak Test	QWI-QC-03	General Insp. Level- I, AQL- 1.5	Leak proof
Autoclaving	QWI-QC-07	General Insp. Level- I, AQL- 1.5	Autoclavable
Dimensional	QSF-QC-02A	General Insp. Level- I, AQL- 1.5	Within tolerance

Product Specifications

Material:	Bottle:	Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.
	Closure:	Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.
BSE/TSE Statement:	The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.	
Non-Pyrogenic:	Not Applicable.	
DNase & RNase Free:	Not Applicable.	
Cytotoxicity:	Not Applicable.	
Latex Free Statement:	No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.	

Signed:



Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: February 22, 2019